

# Cryo clor 0.3



Spray disinfectant for cryostat

**IVD** In-vitro diagnostic medical device  
IVD in **Class A**, Reg. UE 2017/746

Catalog number	Unit size	UDI-DI	BASIC UDI
05-9802	125 ml	08033976233068	080339762W01030799Y5

**Packaging** Primary container: white PE bottle with spray and protective cap. Useful capacity 125 ml.  
Secondary container: carton box.

Wear, water, alcohol and solvents resistant PVC label. Scratchproof ink resistant to water and alcohol.

**Application** Spray disinfectant for cryostat.

Components	CAS	CE	Index
Hydroalcoholic mixture with 0,3% chlorexydine	-	-	-

**Functional properties**

- Cryo clor 0,3 acts in 15 minutes.
- It works at -20°C; it doesn't freeze and it doesn't leave residue.
- Easy to use thanks to spray bottle.
- It makes easier to clean the inside of cryostat.
- Effective against bacteria, fungi, and some viruses (Hepatitis B, Poliomyelitis, Herpes simplex, HIV).
- It can be used as general disinfectant for surfaces in laboratory.

**Warning and precaution** The product is intended for professional laboratory use for healthcare professionals. Carefully read the information on the label (danger symbols, risk and safety phrases) and always consult the safety data sheet. Do not use if the primary container is damaged. In the event of a serious accident, we recommended that you immediately inform Bio-Optica Milano S.p.A and the competent authorities.

**Storage** Store the preparation at 15-25°C. Keep the containers tightly closed.

**Stability** After the first opening, the product is reusable until the expiry date, if correctly stored. Product validity: 2 years.

**Disposal** Hazardous preparation: observe all state and local environmental regulations regarding waste disposal.

REVISION n°	REASON	REVISION DATE
001	Regulation adjustment UE 2017/746 - IVDR	16/05/2022